SECTION - C

TECHNICAL SPECIFICATIONS OF STORES AND DRAWINGS.

<u>Technical Specifications / Compliance Statement for Technical specification of Turbo</u> <u>Molecular Pump (TMP)</u>

Vendor must filled and submit the compliance statement with official seal and signature with offer.

Sr.	Specifications	IPR Requirements	Vendor
No.			Specification
1	Size of gas intake port/ flange of the	DN 200 ISO CF or	
	TMP	Nearest standard size as per vendor	
2	Required pumping Speed	≥7001/s	
	(Nitrogen/Helium)		
4	Ultimate vacuum with baking	~1.0 X 10 ⁻⁸ mbar	
5	Mounting Orientation	Vertical/ Horizontal installation should be	
		possible	
6.	Main Bearing Type	Hybrid Bearing (Magnetic bearing on High	
	0 71	vacuum side and Oil lubricated Ceramic	
		Bearing on fore vacuum side)	
7	Cooling	Air / Water	
8	Baking Temp	100° C (This value is for body	
		temperature of TMP and temperature at	
		pumping port)	
9	Sound level	~50 dB (at a distance of 1 m from TMP	
		Unit)	
10	Splinter Shield	To be provided by vendor	
11	Heating Jacket	To be provided by vendor	
12	Suitable Backing pump	suitable baking pump should be provided	
	Sanasie Baennig pamp	by vendor	
13	Gate Valve	Suitable bellow sealed gate valve of size	
10		DN200, ISO CF, electro-pneumatically	
		actuated, should be provided by vendor (As	
		per specification, included in this document)	
		per specification, merudeu în uns document)	

Electronic Drive / Control unit

Sr. No.	Specifications	Requirements	Vendor Specification
1	Control unit	Display and Control unit suitable forabove pump (Optional: provision of port for interfacing with external PLC should be included and for this if price is different, its price should be separately quoted)	
2	Available electrical power at IPR	440 V \pm 10%, 50 Hz \pm 5Hz , 3 phase and 230 V \pm 10%, 50 Hz \pm 5Hz , single phase.	
3	Interlock provision for failure of backing Pump and venting valve of the pumping Unit	To be provided by Vendor	
4	Control logic for automatic Gate	To be provided by Vendor	

	valve Opening & with speed of TM level in the applie	MP and			
5.	Required conr supply cables	necting	power	~3 m (to be provided by vendor)	

Specification for UHV compatible, electro-pneumatically operated gate valve

Sr. No.	Specifications	Requirements	Vendor Specification
1.	Size of Flange	DN 200 ISO CF	
2.	Orientation	Any	
3.	Actuator	Pneumatic double acting cylinder	
4.	Shaft Sealing	Bellow shaft seal	
5.	Material of main parts	 (a) Valve body and moving mechanism :SS304 (b) Bellow: 304L (c) Gate Plate: Aluminium alloy or suitable Stainless Steel 	
6.	Sealing material	(a) Bonnet: Metal(b) Gate: Viton	
7.	Helium leak tightness	 (a) Body: ≤5×10E-10 mbar l/sec (b) Valve Seat : ≤1×10E-9 mbar l/sec 	
8.	Operating Pressure	(c) 1100 mbar to $\leq 1 \times 10E-9$ mbar	
9.	Available compressed air pressure at IPR for electropnematic	4-5 bar(absolute)	

IPR	Vendor Specification
Accessories: - 1 set of suitable and essentially required spare parts should be quoted along with your offer. Features and specifications	
of these spares parts should include by vendor	
Installation and Acceptance Tests: - The final acceptance will be given after successful operation and achieving the required performance of the TMP unit. Installation will be done by vendor with the vacuum chamber/system available at IPR after which vendor will carry out	
acceptance tests at IPR for following parameters.	
1. Specified ultimate pressure at the suction port of pump	
2. Functioning of control logic and safe operation of TMP in nominal and fault conditions	
3. Noise level should not be more then as per the vendor's specification.	
4. Leak tightness of housing and seat of gate valve should not be more then specified by the vendor.	
5. Maximum and minimum differential pressure at which gate valve can open and it should be as per vendor's specification.	

Documents to be provided to IPR

Sr. No.	IPR	Vendor
		Specification
		Page 2 of 3

1	The offer must be submitted with all necessary technical details, literature, user manual and related documents to the corresponding model	
2	The offer must have a document giving details of spec and its values item wise similar to the format of this document for the offered product. This document is necessary even though vendor is providing literature, brochure and manuals of the product.	
3	Detailed information on operation, maintenance and troubleshooting should be provided.	
General	Terms:	
	Vendor has to install the pumping system and gate valve with the vacuum chamber available at IPR site for acceptance tests and carryout the different tests required for acceptance at IPR.	

Sr. No.	IPR	Vendor Specification
<u>1.</u>	Spare parts: Vendor should provide required spare parts and consumables (like gaskets, lubricating oils etc., for total operation of about 7500 hours distributed over five years. Maximum duration of operation continuously could be for about one week. Separate price list for each spare part for TMP and gate valve should be provided along with the quotation.	
<u>2.</u>	Warrantee / Guarantee : The vendor must guarantee/warrantee the equipment, supplied for a period of one years starting from the date of acceptance by IPR.	

Authorized Signatory

Official Seal

Date :