

This file has been cleaned of potential threats.

To view the reconstructed contents, please SCROLL DOWN to next page.



प्लाज्मा अनुसंधान संस्थान
INSTITUTE FOR PLASMA RESEARCH
 परमाणु ऊर्जा विभाग, भारत सरकार का एक सहायता प्राप्त संस्थान
**An Aided Institute of Department of Atomic Energy,
 Government of India**



इन्दिरा पुल के पास, भट, गांधीनगर - 382 428 भारत
 दूरभाष: (079) 2396 2020/2021/2028
 फैक्स: 91-079-23962277
 वेब: www.ipr.res.in

NEAR INDIRA BRIDGE, BHAT
 DIST. GANDHINAGAR - 382 428 (INDIA)
 Phone: (079) 2396 2000/2026/2332
 Fax : 91-079-23962277
 Web : www.ipr.res.in

ENQUIRY

ENQUIRY NO : IPR/EQF/18-19/068
 Date : 12-06-2018

Due on : 12-07-2018 by 1:00 PM IST

Please send your offer in sealed envelope specifying Enquiry No, Date & Due Date, ALONG WITH your credentials for the following items which we are interested to import directly against Foreign Trade Policy 2015-2020.

Important Note:

Please note that e-mail quotations are not acceptable however you may send your queries (if any) to importpurchase@ipr.res.in

Please ensure your sealed quotation reaches this office not later than above mentioned due date and time.

Kindly go through the following documents properly before quoting which are available on the IPR web portal i.e., http://www.ipr.res.in/documents/tender_terms.html / attached herewith.

- 1) Instructions to the bidders & Terms and conditions (refer Form No: IPR-FP-01.V3)
- 2) Bidding format

GST for Goods and Services (IGST/CGST/SGST TAX BENEFITS): Please refer clause no: 14 of Form No: IPR-FP-01.V3

QUOTATION SHOULD BE ADDRESSED TO PURCHASE OFFICER ONLY

Sr No	Description	Quantity
1	UHV hot cathod ionisation gauge as per specification	1.0 Nos.

Note: Please quote with complete technical details (Technical compliance sheet and product data sheet).

Encl: Refer attached sheet for detailed technical specification.

Sd/-

Mr. D. Ramesh
 Purchase Officer-II

Information to Vendors: We are working towards a single platform for our future requirement. Hence, please refer IPR website i.e., <http://www.ipr.res.in/documents/tenderseng.html> for our future requirement.

Specification for hot cathode UHV gauge

SR. No.	IPR Specification		Vendor specification
1	Measurement range	$\leq 2 \times 10^{-12} - \geq 1 \times 10^{-4}$ mbar	
2	Xray limit	$\leq 1 \times 10^{-12}$ mbar	
3	Bake able temperature w/o cable	At least up to 250° C.	
4	Degassing	Electron bombardment	
5	Vacuum flange type	DN 40 CF	
6	Cathode filament type	Iridium with yttric oxide coating	
7	Display	Digital exponential display for pressure measurement.	
8	Interface	RS232/RS485	
9	Analog output	Logarithmic output for full measuring range	
10	Spares and accessories	Power cable, sensor cables etc.	

Date :

Bidder's Sign and Stamp