

This file has been cleaned of potential threats.

To view the reconstructed contents, please SCROLL DOWN to next page.



प्लाज्मा अनुसंधान संस्थान
INSTITUTE FOR PLASMA RESEARCH
 परमाणु ऊर्जा विभाग, भारत सरकार का एक सहायता प्राप्त संस्थान
**An Aided Institute of Department of Atomic Energy,
 Government of India**



इन्दिरा पुल के पास, भट, गांधीनगर - 382 428 भारत
 दूरभाष: (079) 2396 2020/2021/2028
 फैक्स: 91-079-23962277
 वेब: www.ipr.res.in

NEAR INDIRA BRIDGE, BHAT
 DIST. GANDHINAGAR - 382 428 (INDIA)
 Phone: (079) 2396 2000/2026/2332
 Fax : 91-079-23962277
 Web : www.ipr.res.in

ENQUIRY

ENQUIRY NO : IPR/EQF/18-19/086
 Date : 13-07-2018

Due on : 16-08-2018 by 1:00 PM IST

Please send your offer in sealed envelope specifying Enquiry No, Date & Due Date, ALONG WITH your credentials for the following items which we are interested to import directly against Foreign Trade Policy 2015-2020.

Important Note:

Please note that e-mail quotations are not acceptable however you may send your queries (if any) to importpurchase@ipr.res.in

Please ensure your sealed quotation reaches this office not later than above mentioned due date and time.

Kindly go through the following documents properly before quoting which are available on the IPR web portal i.e., http://www.ipr.res.in/documents/tender_terms.html / attached herewith.

- 1) Instructions to the bidders & Terms and conditions (refer Form No: IPR-FP-01.V3)
- 2) Bidding format

GST for Goods and Services (IGST/CGST/SGST TAX BENEFITS): Please refer **clause no: 14** of Form No: IPR-FP-01.V3

QUOTATION SHOULD BE ADDRESSED TO PURCHASE OFFICER ONLY

Sr No	Description	Quantity
1	100 kV -High Voltage Divider/ Passive Probe as per attached specification	3.0 Nos.

Note: Please quote with complete technical details (Technical compliance sheet and product data sheet).
 Payment will be made as follows :-
 - Through Letter of Credit :- The L/C will be established for 90% value of purchase order and the same will be paid against presentation of shipping documents.
 - Through Wire Transfer : 10% payment will be paid through wire transfer after receipt and final acceptance of the item/s at IPR site and on submission of performance Bank Guarantee for 10% of the order value, valid throughout warranty period.

Encl: Refer attached sheet for detailed technical specification.

Sd/-

Mr. D. Ramesh
Purchase Officer-II

Information to Vendors: We are working towards a single platform for our future requirement. Hence, please refer IPR website i.e, <http://www.ipr.res.in/documents/tendersenq.html> for our future requirement.

Specification for High Voltage Divider/ Passive Probe

Parameter	Value
DC voltage	100 kV
Divider Ratio	10,000:1 (on 1 M Ω terminating impedance)
Accuracy DC @ 1 MHz	$\pm 0.5\%$ or better $\pm 3\%$ or better
Output terminal on divider (LV side)	BNC type
Output cable (shielded) with BNC connectors at both ends	2 meters (50 Ohm impedance)
Additional Requirements : (1) Intended for Indoor application. (2) Appropriate corona rings should be part of Divider/ Probe. (3) In case of order, manufacturer has to provide factory test report and conformity certificate according to IEC 61010-1 or equivalent before dispatch.	

Note: Catalogue for the offered divider/probe and General Arrangement Drawings should be provided along with offer.