

This file has been cleaned of potential threats.

To view the reconstructed contents, please SCROLL DOWN to next page.



प्लाज्मा अनुसंधान संस्थान  
**INSTITUTE FOR PLASMA RESEARCH**  
 परमाणु ऊर्जा विभाग, भारत सरकार का एक सहायता प्राप्त संस्थान  
**An Aided Institute of Department of Atomic Energy,  
 Government of India**



इन्दिरा पुल के पास, भट, गांधीनगर - 382 428 भारत  
 दूरभाष: (079) 2396 2020/2021/2028  
 फैक्स: 91-079-23962277  
 वेब: [www.ipr.res.in](http://www.ipr.res.in)

NEAR INDIRA BRIDGE, BHAT  
 DIST. GANDHINAGAR - 382 428 (INDIA)  
 Phone: (079) 2396 2000/2026/2332  
 Fax : 91-079-23962277  
 Web : [www.ipr.res.in](http://www.ipr.res.in)

## ENQUIRY

ENQUIRY NO : IPR/EQF/18-19/088  
 Date : 16-07-2018

**Due on : 23-08-2018 by 1:00 PM IST**

Please send your offer in sealed envelope specifying Enquiry No, Date & Due Date, ALONG WITH your credentials for the following items which we are interested to import directly against Foreign Trade Policy 2015-2020.

### Important Note:

Please note that e-mail quotations are not acceptable however you may send your queries (if any) to [importpurchase@ipr.res.in](mailto:importpurchase@ipr.res.in)

Please ensure your sealed quotation reaches this office not later than above mentioned due date and time.

Kindly go through the following documents properly before quoting which are available on the IPR web portal i.e., [http://www.ipr.res.in/documents/tender\\_terms.html](http://www.ipr.res.in/documents/tender_terms.html) / attached herewith.

- 1) Instructions to the bidders & Terms and conditions (refer Form No: IPR-FP-01.V3)
- 2) Bidding format

**GST for Goods and Services (IGST/CGST/SGST TAX BENEFITS):** Please refer clause no: 14 of Form No: IPR-FP-01.V3

### QUOTATION SHOULD BE ADDRESSED TO PURCHASE OFFICER ONLY

Sr No	Description	Quantity
1	63CF UHV Compatible Re-entrant View Port	2.0 Nos.

Note: Please quote with complete technical details (Technical compliance sheet and product data sheet).

Encl: Refer attached sheet for detailed technical specification.

Sd/-

Mr. D. Ramesh  
 Purchase Officer-II

**Information to Vendors:** We are working towards a single platform for our future requirement. Hence, please refer IPR website i.e., <http://www.ipr.res.in/documents/tenderseng.html> for our future requirement.

## **63CF UHV Compatible Re-entrant View Port**

### **Specification**

Type : 63CF UHV compatible re-entrant view port

Viewable area :  $\geq 35$  mm

Tube length : ~160 mm

Tube inner diameter ~ 48 mm

Viewing window : Quartz or Fused silica

Wavelength range : 200 to 800 nm

Material : SS 304L

**Compliance form****63CF Re-entrant view port**

Sr. No	Particular	IPR Requirement	Vendor Specification
1	Type	63CF UHV compatible re-entrant view port	
2	Viewable area	$\geq 35$ mm	
3	Tube length	~ 160 mm	
4	Tube inner diameter	~ 48 mm	
5	Viewing window	Quartz or Fused silica	
6	Wavelength range	200 - 800 nm	
7	Material	SS 304L	

**Date :-****Bidders sign with official stamp**