

This file has been cleaned of potential threats.

To view the reconstructed contents, please SCROLL DOWN to next page.



प्लाज्मा अनुसंधान संस्थान
INSTITUTE FOR PLASMA RESEARCH
परमाणु ऊर्जा विभाग, भारत सरकार का एक सहायता प्राप्त संस्थान
**An Aided Institute of Department of Atomic Energy,
Government of India**



इन्दिरा पुल के पास, भट, गांधीनगर - 382 428 भारत
दूरभाष: (079) 2396 2020/2021/2028
फैक्स: 91-079-23962277
वेब: www.ipr.res.in

NEAR INDIRA BRIDGE, BHAT
DIST. GANDHINAGAR - 382 428 (INDIA)
Phone: (079) 2396 2000/2026/2332
Fax : 91-079-23962277
Web : www.ipr.res.in

ENQUIRY

ENQUIRY NO : IPR/EQF/18-19/132
Date : 26-09-2018

Due on : 01-11-2018 by 1:00 PM IST

Please send your offer in sealed envelope specifying Enquiry No, Date & Due Date, ALONG WITH your credentials for the following items which we are interested to import directly against Foreign Trade Policy 2015-2020.

Important Note:

Please note that e-mail quotations are not acceptable however you may send your queries (if any) to importpurchase@ipr.res.in

Please ensure your sealed quotation reaches this office not later than above mentioned due date and time.

Kindly go through the following documents properly before quoting which are available on the IPR web portal i.e., http://www.ipr.res.in/documents/tender_terms.html / attached herewith.

- 1) Instructions to the bidders & Terms and conditions (refer Form No: **IPR-FP-01.V3**)
- 2) Bidding format

GST for Goods and Services (IGST/CGST/SGST TAX BENEFITS): Please refer **clause no: 14** of Form No: **IPR-FP-01.V3**

QUOTATION SHOULD BE ADDRESSED TO PURCHASE OFFICER ONLY

Sr No	Description	Quantity
1	Visible band pass filter sets as per the specification sheet attached	2.0 Sets

Note: Please quote with complete technical details (Technical compliance sheet and product data sheet).

Encl: Other details are as per attached specification sheet.

Sd/-

Mr. D. Ramesh
Purchase Officer-II

Information to Vendors: We are working towards a single platform for our future requirement. Hence, please refer IPR website i.e,

<http://www.ipr.res.in/documents/tendersenq.html> for our future requirement.

Technical specification for visible band pass filter set

IPR specification	
Filter type:	Band Pass
Substrate material:	Schott Borofloat/ Soda Lime/Fused Silica/NBK-7
Diameter:	1 inch/25mm
Thickness:	≤ 6mm
Clear aperture:	≥85%
Surface quality:	80-50 or better
Central wavelengths with tolerance:	350±2nm, 400 ±2nm, 450 ±2nm, 500 ±2nm, 550±2nm, 600±2nm, 650±2nm, 700±2nm, 750±2nm, 800±2nm
FWHM:	10±2nm
Out of band transmission:	< 0.01% from 200 nm to 3μm
Edge Treatment:	Mounted in black anodized aluminum ring
Edge marking:	An arrow indicating the transmission direction
Operating temperature:	-50°C to +80°C
Nos of band pass filter	2 in each

Note: In the compliance table pl. write the exact specification.

Pl. do not write 'confirm', 'yes' or 'OK' etc.

Technical compliance form

IPR specification	Compliance from Vendor
Filter type: Band Pass	
Substrate material: Schott Borofloat/ Soda Lime/Fused Silica/NBK-7	
Diameter: 1 inch/25mm	
Thickness: ≤ 6 mm	
Clear aperture: $\geq 85\%$	
Surface quality: 80-50 or better	
Central wavelengths with tolerance: 350 ± 2 nm, 400 ± 2 nm, 450 ± 2 nm, 500 ± 2 nm, 550 ± 2 nm, 600 ± 2 nm, 650 ± 2 nm, 700 ± 2 nm, 750 ± 2 nm, 800 ± 2 nm	
FWHM: 10 ± 2 nm	
Out of band transmission: $< 0.01\%$ from 200 nm to 3 μ m	
Edge Treatment: Mounted in black anodized aluminum ring	
Edge marking: An arrow indicating the transmission direction	
Operating temperature: -50 $^{\circ}$ C to +80 $^{\circ}$ C	
Nos of band pass filter : 2 in each	

Date: _____

Bidder's stamp and Sign: _____