

This file has been cleaned of potential threats.

To view the reconstructed contents, please SCROLL DOWN to next page.



प्लाज्मा अनुसंधान संस्थान
INSTITUTE FOR PLASMA RESEARCH
परमाणु ऊर्जा विभाग, भारत सरकार का एक सहायता प्राप्त संस्थान
**An Aided Institute of Department of Atomic Energy,
Government of India**



इन्दिरा पुल के पास, भट, गांधीनगर - 382 428 भारत
दूरभाष: (079) 2396 2020/2021/2028
फैक्स: 91-079-23962277
वेब: www.ipr.res.in

NEAR INDIRA BRIDGE, BHAT
DIST. GANDHINAGAR - 382 428 (INDIA)
Phone: (079) 2396 2000/2026/2332
Fax : 91-079-23962277
Web : www.ipr.res.in

ENQUIRY

ENQUIRY NO : IPR/EQF/18-19/160
Date : 13-11-2018

Due on : 20-12-2018 by 1:00 PM IST

Please send your offer in sealed envelope specifying Enquiry No, Date & Due Date, ALONG WITH your credentials for the following items which we are interested to import directly against Foreign Trade Policy 2015-2020.

Important Note:

Please note that e-mail quotations are not acceptable however you may send your queries (if any) to importpurchase@ipr.res.in

Please ensure your sealed quotation reaches this office not later than above mentioned due date and time.

Kindly go through the following documents properly before quoting which are available on the IPR web portal i.e., http://www.ipr.res.in/documents/tender_terms.html / attached herewith.

- 1) Instructions to the bidders & Terms and conditions (refer Form No: IPR-FP-01.V3)
- 2) Bidding format

GST for Goods and Services (IGST/CGST/SGST TAX BENEFITS): Please refer clause no: 14 of Form No: IPR-FP-01.V3

QUOTATION SHOULD BE ADDRESSED TO PURCHASE OFFICER ONLY

Sr No	Description	Quantity
1	Re-entrant viewport as per the specification sheet	2.0 Nos.

Note: 1. Please quote with complete technical details (Technical compliance sheet and product data sheet).
2. TDS as per CGST Act: As per provisions of section No. 51 of the CGST Act 2017, TDS @2% (IGST 2% or CGST 1% and SGST 1%) will be deducted while making payment to the suppliers where total value of orders/contracts/work orders exceeds Rs. 2.5 lakhs, in the event of order in Indian Rupees. Necessary TDS Certificate will be issued to the supplier after TDS deduction.

Encl: Other details are as per attached specification sheet.

Sd/-

Mr. D. Ramesh

Information to Vendors: We are working towards a single platform for our future requirement. Hence, please refer IPR website i.e, <http://www.ipr.res.in/documents/tendersenq.html> for our future requirement.

63CF UHV Compatible Re-entrant View Port

Spécification Sheet

Quantity : 02 Nos.

Type : 63CF UHV compatible re-entrant view port

Viewable area: ≥ 35 mm

Tube length: ~160 mm

Tube inner diameter ~ 48 mm

Viewing window: Quartz or Fused silica

Wavelength range: 200 to 800 nm

Material : Stainless Steel 304L

Vacuum Compatibility: UHV (ultra high vacuum) compatible.

Baking Temperature: the assembly should withstand 200 degree celsius temperature for 05 hours.

Bidder's stamp and Sign:

Date:

Compliance Sheet

Particulars	IPR Specification	Vendor's Specification
Quantity	02 Nos.	
Flange type	63CF UHV compatible re-entrant view port	
Viewable area:	~ 35 mm	
Tube length:	~160 mm	
Tube inner-diameter	~ 48 mm	
Viewing window:	Quartz or Fused silica	
Wavelength range:	200 to 800 nm	
Material:	Stainless Steel 304L	
Vacuum Compatibility:	UHV (ultra-high vacuum) compatible.	
Baking Temperature:	The assembly should withstand 200 °C temperature for 05 hours.	

Bidder's Sign with Official Stamp:

Date: