

This file has been cleaned of potential threats.

To view the reconstructed contents, please SCROLL DOWN to next page.



प्लाज्मा अनुसंधान संस्थान
INSTITUTE FOR PLASMA RESEARCH
परमाणु ऊर्जा विभाग, भारत सरकार का एक सहायता प्राप्त संस्थान
**An Aided Institute of Department of Atomic Energy,
Government of India**



इन्दिरा पुल के पास, भट, गांधीनगर - 382 428 भारत
दूरभाष: (079) 2396 2020/2021/2028
फैक्स: 91-079-23962277
वेब: www.ipr.res.in

NEAR INDIRA BRIDGE, BHAT
DIST. GANDHINAGAR - 382 428 (INDIA)
Phone: (079) 2396 2000/2026/2332
Fax : 91-079-23962277
Web : www.ipr.res.in

ENQUIRY

ENQUIRY NO : IPR/EQF/18-19/166
Date : 28-11-2018

Due on : 03-01-2019 by 1:00 PM IST

Please send your offer in sealed envelope specifying Enquiry No, Date & Due Date, ALONG WITH your credentials for the following items which we are interested to import directly against Foreign Trade Policy 2015-2020.

Important Note:

Please note that e-mail quotations are not acceptable however you may send your queries (if any) to importpurchase@ipr.res.in

Please ensure your sealed quotation reaches this office not later than above mentioned due date and time.

Kindly go through the following documents properly before quoting which are available on the IPR web portal i.e., http://www.ipr.res.in/documents/tender_terms.html / attached herewith.

- 1) Instructions to the bidders & Terms and conditions (refer Form No: **IPR-FP-01.V3**)
- 2) Bidding format

GST for Goods and Services (IGST/CGST/SGST TAX BENEFITS): Please refer **clause no: 14** of Form No: **IPR-FP-01.V3**

QUOTATION SHOULD BE ADDRESSED TO PURCHASE OFFICER ONLY

Sr No	Description	Quantity
1	Line Camera with USB interface and external trigger. Other specifications as attached.	7.0 Nos.

Note: 1. Please quote with complete technical details (Technical compliance sheet and product data sheet).
2. TDS as per CGST Act: As per provisions of section No. 51 of the CGST Act 2017, TDS @2% (IGST 2% or CGST 1% and SGST 1%) will be deducted while making payment to the suppliers where total value of orders/contracts/work orders exceeds Rs. 2.5 lakhs, in the event of order in Indian Rupees. Necessary TDS Certificate will be issued to the supplier after TDS deduction.

Encl: Other details are as per attached specification sheet.

Sd/-

Mr. D. Ramesh
Purchase Officer-II

Information to Vendors: We are working towards a single platform for our future requirement. Hence, please refer IPR website i.e, <http://www.ipr.res.in/documents/tendersenq.html> for our future requirement.

Technical specifications for Line scan Camera with USB interface

Quantity required: 7 cameras

IPR Parameter		
#	Technical Parameter	Required Value/ type
1.	Sensor Type	CCD
2.	No. Of pixels	2048 or better (linear array)
3.	Active Length of Sensor	25 mm or better
4.	Spectral range	350 to 1000 nm
5.	ADC resolution	12 bit or better
6.	Integration time/Exposure time	300 uSec or better
7.	Frame rate	250 fps or better
8.	Pixel output clock	1 MHz or better
9.	LabVIEW support	Required
10.	External trigger support	Required
11.	Interface	USB 2.0
12.	Power bus	USB powered
13.	Operating system support	Windows 7

Compliance sheet

IPR Parameter			Vendor Parameter
#	Technical Parameter	Required Value/ type	
1.	Sensor Type	CCD	
2.	No. Of pixels	2048 or better (linear array)	
3.	Active Length of Sensor	25 mm or better	
4.	Spectral range	350 to 1000 nm	
5.	ADC resolution	12 bit or better	
6.	Integration time/Exposure time	300 uSec or better	
7.	Frame rate	250 fps or better	
8.	Pixel output clock	1 MHz or better	
9.	LabVIEW support	Required	
10.	External trigger support	Required	
11.	Interface	USB 2.0	
12.	Power bus	USB powered	
13.	Operating system support	Windows 7	

Bidder's stamp and Sign:

Date: