

This file has been cleaned of potential threats.

To view the reconstructed contents, please SCROLL DOWN to next page.



प्लाज्मा अनुसंधान संस्थान
INSTITUTE FOR PLASMA RESEARCH
परमाणु ऊर्जा विभाग, भारत सरकार का एक सहायता प्राप्त संस्थान
**An Aided Institute of Department of Atomic Energy,
Government of India**



इन्दिरा पुल के पास, भट, गांधीनगर - 382 428 भारत
दूरभाष: (079) 2396 2020/2021/2028
फैक्स: 91-079-23962277
वेब: www.ipr.res.in

NEAR INDIRA BRIDGE, BHAT
DIST. GANDHINAGAR - 382 428 (INDIA)
Phone: (079) 2396 2000/2026/2332
Fax : 91-079-23962277
Web : www.ipr.res.in

ENQUIRY

ENQUIRY NO : IPR/EQF/18-19/188
Date : 16-01-2019

Due on : 28-03-2019 by 1:00 PM IST

Reminder-1 Dt: 22-02-2019

Please send your offer in sealed envelope specifying Enquiry No, Date & Due Date, ALONG WITH your credentials for the following items which we are interested to import directly against Foreign Trade Policy 2015-2020.

Important Note:

Please note that e-mail quotations are not acceptable however you may send your queries (if any) to importpurchase@ipr.res.in

Please ensure your sealed quotation reaches this office not later than above mentioned due date and time.

Kindly go through the following documents properly before quoting which are available on the IPR web portal i.e., http://www.ipr.res.in/documents/tender_terms.html / attached herewith.

- 1) Instructions to the bidders & Terms and conditions (refer Form No: **IPR-FP-01.V3**)
- 2) Bidding format

GST for Goods and Services (IGST/CGST/SGST TAX BENEFITS): Please refer clause no: 14 of Form No: **IPR-FP-01.V3**

QUOTATION SHOULD BE ADDRESSED TO PURCHASE OFFICER ONLY

Sr No	Description	Quantity
1	Surface Activity Monitor Including Calibration Source	1.0 Nos.

Note: 1. Please quote with complete technical details (Technical compliance sheet and product data sheet).
2. TDS as per CGST Act: As per provisions of section No. 51 of the CGST Act 2017, TDS @2% (IGST 2% or CGST 1% and SGST 1%) will be deducted while making payment to the suppliers where total value of orders/contracts/work orders exceeds Rs. 2.5 lakhs, in the event of order in Indian Rupees. Necessary TDS Certificate will be issued to the supplier after TDS deduction.

Encl: As per attached.

Sd/-
Mr. D. Ramesh
Purchase Officer-II

Information to Vendors: We are working towards a single platform for our future requirement. Hence, please refer IPR website i.e, <http://www.ipr.res.in/documents/tendersenq.html> for our future requirement.

Technical Compliance of Surface Activity Monitor for Tritium

Surface Activity Monitor			
Sr. No.	Particulars	IPR Specification	Vendor's Specification
1	Detection Range	0-1600 nci/100 cm ²	
2	Detection Sensitivity	+/- 10 nCi/100 cm ² or better	
3	Response Time	≤ 2 minutes	
4	Measurement Area	100 cm ²	
5	Display	LCD Display	
6	Process/Ambient Temperature	15°C to 50°C	
7	Tritium Wetted Parts	316L Stainless Steel, Teflon	
8	Weight	maximum 5 kg	
9	Measurement Units	nCi/100 cm ² or Bq/cm ²	
10	Alarm level	setting Full range	
Calibration Source			
11	A calibrated tritium emitter source of company recommended activity should be provided with the instrument.		
Acceptance Criteria			
12	Calibration certificate should be submitted along with shipment.		

- For Sr. no. 1 to 10: Please provided the technical value/data. Please do not fill with words like “yes”, “Comply” or any other vague words.
- For Sr. no. 11 to 12: Mention whether it will be provided/agreed.

Bidder's sign with official stamp

Date: