

This file has been cleaned of potential threats.

To view the reconstructed contents, please SCROLL DOWN to next page.



प्लाज़्मा अनुसंधान संस्थान  
INSTITUTE FOR PLASMA RESEARCH  
परमाणु ऊर्जा विभाग, भारत सरकार का एक सहायता प्राप्त  
संस्थान



An Aided Institute of Department of Atomic Energy,  
Government of India

इन्दिरा पुल के पास, भट, गांधीनगर - 382 428 भारत  
दूरभाष: (079) 2396 2020/2021/2028  
फैक्स: 91-079-23962277  
वेब: [www.ipr.res.in](http://www.ipr.res.in)

NEAR INDIRA BRIDGE, BHAT  
DIST. GANDHINAGAR - 382 428 (INDIA)  
Phone: (079) 2396 2000/2026/2332  
Fax : 91-079-23962277  
Web : [www.ipr.res.in](http://www.ipr.res.in)

### ENQUIRY

ENQUIRY NO : IPR/EQF/18-19/192  
Date : 28-01-2019

**Due on : 07-03-2019 by 1:00 PM IST**

Please send your offer in sealed envelope specifying Enquiry No, Date & Due Date, ALONG WITH your credentials for the following items which we are interested to import directly against Foreign Trade Policy 2015-2020.

#### Important Note:

Please note that e-mail quotations are not acceptable however you may send your queries (if any) to [importpurchase@ipr.res.in](mailto:importpurchase@ipr.res.in)

Please ensure your sealed quotation reaches this office not later than above mentioned due date and time.

Kindly go through the following documents properly before quoting which are available on the IPR web portal i.e., [http://www.ipr.res.in/documents/tender\\_terms.html](http://www.ipr.res.in/documents/tender_terms.html) / attached herewith.

- 1) Instructions to the bidders & Terms and conditions (refer Form No: **IPR-FP-01.V3**)
- 2) Bidding format

**GST for Goods and Services (IGST/CGST/SGST TAX BENEFITS):** Please refer **clause no: 14** of Form No: **IPR-FP-01.V3**

#### QUOTATION SHOULD BE ADDRESSED TO PURCHASE OFFICER ONLY

Sr No	Description	Quantity
1	63CF standard View Port Glass: Kodial Sealing Transition: Kovar/Steel/Alumina Flange: SS304	4.0 Nos.

	Baking Temperature: 250 degree C Transmission: Visible Spectrum	
2	35/40CF standard View Port Glass: Kodial Sealing Transition: Kovar/Steel/Alumina Flange: SS304 Baking Temperature: 250 degree C Transmission: Visible Spectrum	4.0 Nos.

Note: 1. Please quote with complete technical details (Technical compliance sheet and product data sheet).  
2. TDS as per CGST Act: As per provisions of section No. 51 of the CGST Act 2017, TDS @2% (IGST 2% or CGST 1% and SGST 1%) will be deducted while making payment to the suppliers where total value of orders/contracts/work orders exceeds Rs. 2.5 lakhs, in the event of order in Indian Rupees. Necessary TDS Certificate will be issued to the supplier after TDS deduction.

Encl: Refer attached sheet for detailed technical specification.

Sd/-

Mr. D. Ramesh  
Purchase Officer-II

**Information to Vendors:** We are working towards a single platform for our future requirement. Hence, please refer IPR website i.e, <http://www.ipr.res.in/documents/tendersenq.html> for our future requirement.

## **Technical Specification**

### **1. Standard View Port: 63CF**

- a. Vacuum Flange: 63 CF
- b. View Port Glass: Kodial
- c. Transition Seal: Kovar/ Steel/ Alumina
- d. Flange: SS304
- e. Baking Temperature: 250 Degree C
- f. Transmission: Visible Spectrum

### **2. Standard View Port: 35/40CF**

- a. Vacuum Flange: 35/40 CF
- b. View Port Glass: Kodial
- c. Transition Seal: Kovar/ Steel/ Alumina
- d. Flange: SS304
- e. Baking Temperature: 250 Degree C
- f. Transmission: Visible Spectrum

## Compliance Statement

### 1. Standard View Port: 63CF

<b>IPR Technical Specification</b>	<b>Vendor's Compliance</b>
Vacuum Flange: 63 CF	
View Port Glass: Kodial	
Transition Seal: Kovar/ Steel/ Alumina	
Flange: SS304	
Baking Temperature: 250 Degree C	
Transmission: Visible Spectrum	

### 2. Standard View Port: 35/40CF

<b>IPR Technical Specification</b>	<b>Vendor's Compliance</b>
Vacuum Flange: 35/40 CF	
View Port Glass: Kodial	
Transition Seal: Kovar/ Steel/ Alumina	
Flange: SS304	
Baking Temperature: 250 Degree C	
Transmission: Visible Spectrum	

Bidder's Official Stamp and Sign

Date: