

This file has been cleaned of potential threats.

To view the reconstructed contents, please SCROLL DOWN to next page.



प्लाज्मा अनुसंधान संस्थान
INSTITUTE FOR PLASMA RESEARCH
 परमाणु ऊर्जा विभाग, भारत सरकार का एक सहायता प्राप्त संस्थान
**An Aided Institute of Department of Atomic Energy,
 Government of India**



इन्दिरा पुल के पास, भट, गांधीनगर - 382 428 भारत
 दूरभाष: (079) 2396 2020/2021/2028
 फैक्स: 91-079-23962277
 वेब: www.ipr.res.in

NEAR INDIRA BRIDGE, BHAT
 DIST. GANDHINAGAR - 382 428 (INDIA)
 Phone: (079) 2396 2000/2026/2332
 Fax : 91-079-23962277
 Web : www.ipr.res.in

ENQUIRY

Office Copy ENQUIRY NO : IPR/EQF/18-19/195
 Date : 30-01-2019
 Type : General

Due on : 07-03-2019 by 1:00 PM IST

Please send your offer in sealed envelope specifying Enquiry No, Date & Due Date, ALONG WITH your credentials for the following items which we are interested to import directly against Foreign Trade Policy 2015-2020.

Important Note:

Please note that e-mail quotations are not acceptable however you may send your queries (if any) to importpurchase@ipr.res.in

Please ensure your sealed quotation reaches this office not later than above mentioned due date and time.

Kindly go through the following documents properly before quoting which are available on the IPR web portal i.e., http://www.ipr.res.in/documents/tender_terms.html / attached herewith.

- 1) Instructions to the bidders & Terms and conditions (refer Form No: IPR-FP-01.V3)
- 2) Bidding format

GST for Goods and Services (IGST/CGST/SGST TAX BENEFITS): Please refer clause no: 14 of Form No: IPR-FP-01.V3

QUOTATION SHOULD BE ADDRESSED TO PURCHASE OFFICER ONLY

Sr No	Description	Quantity
1	Supply of Ionization vacuum gauge as per attached specifications sheet	4.0 Nos.

Note:

1. Please quote with complete technical details (Technical compliance sheet and product data sheet).
2. TDS as per CGST Act: As per provisions of section No. 51 of the CGST Act 2017, TDS @2% (IGST 2% or CGST 1% and SGST 1%) will be deducted while making payment to the suppliers where total value of orders/contracts/work orders exceeds Rs. 2.5 lakhs, in the event of order in Indian Rupees. Necessary TDS Certificate will be issued to the supplier after TDS deduction.
3. Unsigned quotations will not be considered. Quotation should be submitted duly signed with official seal on each and every page invariably.

Encl: Refer attached sheet for detailed technical specification.

Mr. D. Ramesh
Purchase Officer-II

Information to Vendors: We are working towards a single platform for our future requirement. Hence, please refer IPR website i.e, <http://www.ipr.res.in/documents/tendersenq.html> for our future requirement.

Specification Sheet for Ionization Vacuum Gauge

Module type	: Built in Sensor, controller and display
Measurement Range	: 1 atmosphere to 1.0×10^{-8} Torr
Display	: Inbuilt Digital display
Analog Output	: Provision for standard analog output for PLC connection required
Overpressure protection	: Gauge turns off at factory set upper pressure limit
Operating temperature	: + 10 °C to + 40 °C
Vacuum Inlet Fitting/flange	: KF 25
Quantity	: 4 No.

Accessories:

Power adaptor	: 230VAC to DC Power adaptor
---------------	------------------------------

Note: Vendors are requested to quote along with all the required accessories.

Offer Compliance:

- i) Vendor shall submit their offer along with the compliances of specifications mentioned above.
- ii) All gauges shall be calibrated at vendor's site and a certificate for the same must be sent along with the delivery of these gauges.

Compliance Form for Supply of Ionization Vacuum Gauges

Sl. No.	Particulars	IPR Requirement	Vendor's Specification
1.	Module type	Built in Sensor, controller and display (Single Integrated unit)	
2.	Measurement Range	1 atmosphere to 1.0×10^{-8} Torr	
3.	Display	Inbuilt Digital display	
4.	Analog Output	Provision for standard analog output for PLC connection required	
5.	Overpressure protection	Gauge turns off at factory set upper pressure limit	
6.	Vacuum Inlet Fitting/flange	KF 25	
7.	Operating temperature	+ 10 °C to + 40 °C	
8.	Quantity	4 Nos.	
9.	Power adaptor	230VAC to DC Power adaptor	
10.	Calibration	Calibration Certificate must be provided by the vendor for all the gauges at the time of delivery	

Bidder's Sign with Official Stamp