

This file has been cleaned of potential threats.

To view the reconstructed contents, please SCROLL DOWN to next page.



प्लाज़्मा अनुसंधान संस्थान
INSTITUTE FOR PLASMA RESEARCH
परमाणु ऊर्जा विभाग, भारत सरकार का एक सहायता
प्राप्त संस्थान



An Aided Institute of Department of Atomic Energy,
Government of India

इन्दिरा पुल के पास, भट, गांधीनगर - 382 428 भारत
दूरभाष: (079) 2396 2020/2021/2028
फैक्स: 91-079-23962277
वेब: www.ipr.res.in

NEAR INDIRA BRIDGE, BHAT
DIST. GANDHINAGAR - 382 428 (INDIA)
Phone: (079) 2396 2000/2026/2332
Fax : 91-079-23962277
Web : www.ipr.res.in

ENQUIRY

ENQUIRY NO : IPR/EQF/18-19/229
Date : 18-03-2019

Due on : 25-04-2019 by 1:00 PM IST

Please send your offer in sealed envelope specifying Enquiry No, Date & Due Date, ALONG WITH your credentials for the following items which we are interested to import directly against Foreign Trade Policy 2015-2020.

Important Note:

Please note that e-mail quotations are not acceptable however you may send your queries (if any) to importpurchase@ipr.res.in

Please ensure your sealed quotation reaches this office not later than above mentioned due date and time.

Kindly go through the following documents properly before quoting which are available on the IPR web portal i.e., http://www.ipr.res.in/documents/tender_terms.html / attached herewith.

- 1) Instructions to the bidders & Terms and conditions (refer Form No: **IPR-FP-01.V3**)
- 2) Bidding format

GST for Goods and Services (IGST/CGST/SGST TAX BENEFITS): Please refer **clause no: 14** of Form No: **IPR-FP-01.V3**

QUOTATION SHOULD BE ADDRESSED TO PURCHASE OFFICER ONLY

Sr No	Description	Quantity
1	Fixed RF Attenuator 20 dB, 100W, N type	2.0 Nos.
2	Fixed RF Attenuator 30 dB, 100W, N type	2.0 Nos.

Note: 1. Please quote with complete technical details (Technical compliance sheet and product data sheet).
2. TDS as per CGST Act: As per provisions of section No. 51 of the CGST Act 2017, TDS @2% (IGST 2% or CGST 1% and SGST 1%) will be deducted while making payment to the suppliers where total value of

orders/contracts/work orders exceeds Rs. 2.5 lakhs, in the event of order in Indian Rupees. Necessary TDS Certificate will be issued to the supplier after TDS deduction.

Encl: refer attached sheet for detailed technical specification

Sd/-

Mr. D. Ramesh
Purchase Officer-II

Information to Vendors: We are working towards a single platform for our future requirement. Hence, please refer IPR website i.e, <http://www.ipr.res.in/documents/tendersenq.html> for our future requirement.

Attenuator's Specifications and compliance form

Attenuators 20 dB

SI No	IPR requirement	Details	Vendor Specifications
1	make & model no		
2	Frequency	DC to 1 GHz or more	
3	Attenuation	20 dB	
4	Accuracy or Flatness	±1 % or less	
5	VSWR	1.4 or less	
6	Input Power	100 W	
7	Reverse Power (if directional)	20 W	
8	Input power peak	1kwatt for 5micro S	
9	Impedance	50 Ohms	
10	Connector	N type male and female	
11	Coolant Method	Heatsink	
	quantity :	2 nos	

Attenuators 30dB

SI No	IPR requirement	Details	Vendor Specifications
1	make & model no		
1	Frequency	DC to 1 GHz or more	
2	Attenuation	30 dB	
3	Accuracy or Flatness	±1 % or less	
4	VSWR	1.4 or less	
5	Input Power	100 W	
6	Reverse Power (if directional)	20 W	
7	Input power peak	1kwatt for 5micro S	
8	Impedance	50 Ohms	
9	Connector	N type male and female	
10	Coolant Method	Heatsink	
	quantity :	2 nos	

Note:- Please fill the technical compliance table along with your offer. Please fill with numbers and statements. Don't just write complied or accepted.

Provide manufacturer datasheet with compliance form.

Bidder' Sign and Stamp

Date