

This file has been cleaned of potential threats.

To view the reconstructed contents, please SCROLL DOWN to next page.



प्लाज्मा अनुसंधान संस्थान
INSTITUTE FOR PLASMA RESEARCH
परमाणु ऊर्जा विभाग, भारत सरकार का एक सहायता प्राप्त संस्थान
**An Aided Institute of Department of Atomic Energy,
Government of India**



इन्दिरा पुल के पास, भट, गांधीनगर - 382 428 भारत
दूरभाष: (079) 2396 2020/2021/2028
फैक्स: 91-079-23962277
वेब: www.ipr.res.in

NEAR INDIRA BRIDGE, BHAT
DIST. GANDHINAGAR - 382 428 (INDIA)
Phone: (079) 2396 2000/2026/2332
Fax : 91-079-23962277
Web : www.ipr.res.in

ENQUIRY

ENQUIRY NO : IPR/EQF/19-20/041
Date : 20-05-2019

Due on : 08-08-2019 by 1:00 PM IST

Reminder-1 Dt: 04-07-2019

Please send your offer in sealed envelope specifying Enquiry No, Date & Due Date, ALONG WITH your credentials for the following items which we are interested to import directly against Foreign Trade Policy 2015-2020.

Important Note:

Please note that e-mail quotations are not acceptable however you may send your queries (if any) to importpurchase@ipr.res.in

Please ensure your sealed quotation reaches this office not later than above mentioned due date and time.

Kindly go through the following documents properly before quoting which are available on the IPR web portal i.e., http://www.ipr.res.in/documents/tender_terms.html / attached herewith.

- 1) Instructions to the bidders & Terms and conditions (refer Form No: **IPR-FP-01.V3**)
- 2) Bidding format

GST for Goods and Services (IGST/CGST/SGST TAX BENEFITS): Please refer clause no: 14 of Form No: **IPR-FP-01.V3**

QUOTATION SHOULD BE ADDRESSED TO PURCHASE OFFICER ONLY

Sr No	Description	Quantity
1	Stop Off Fluid or paste for Brazing filler metal	1.0 Kgs

Note: 1. Please quote with complete technical details (Technical compliance sheet and product data sheet).
2. TDS as per CGST Act: As per provisions of section No. 51 of the CGST Act 2017, TDS @2% (IGST 2% or CGST 1% and SGST 1%) will be deducted while making payment to the suppliers where total value of orders/contracts/work orders exceeds Rs. 2.5 lakhs, in the event of order in Indian Rupees. Necessary TDS Certificate will be issued to the supplier after TDS deduction.

Encl: As per attached

Sd/-
Mr. D. Ramesh
Purchase Officer-II

Information to Vendors: We are working towards a single platform for our future requirement. Hence, please refer IPR website i.e, <http://www.ipr.res.in/documents/tendersenq.html> for our future requirement.

Technical Specification

1	Item	Stop off Fluid or paste
2	Material Composition	Al ₂ O ₃ (≥50% Concentrate by weight) Mixture
3	Application	To apply as a barrier to prevent the unwanted molten flow of filler material during brazing
4	Working Temperature	500 °C to 2000 °C
5	Vacuum Compatibility	It should be compatible furnace brazing with vacuum range of 10 ⁻⁶ mbar to 10 ⁻⁵ mbar
6	Adhesive to Material	After brazing it must be easily removed without leaving any traces behind
7	Quantity	1Kg ± 100 grams

Note: Material should be dispatched along with material test certificate and material safety data sheet.

Technical Compliance

Sr. No.	Item Description	IPR Specification	Vendor Specification
1	Item	Stop off Fluid or paste	
2	Material Composition	Al ₂ O ₃ (≥50% Concentrate by weight) Mixture	
3	Application	To apply as a barrier to prevent the unwanted molten flow of filler material during brazing	
4	Working Temperature	500 °C to 2000 °C	
5	Vacuum Compatibility	It should be compatible furnace brazing with vacuum range of 10 ⁻⁶ mbar to 10 ⁻⁵ mbar	
6	Adhesive to Material	After brazing it must be easily removed without leaving any traces behind	
7	Quantity	1Kg ± 100 grams	
	Note: Material should be dispatched along with material test certificate and material safety data sheet.		

Bidder's Official Stamp and Sign:

Date: