

This file has been cleaned of potential threats.

To view the reconstructed contents, please SCROLL DOWN to next page.



प्लाज्मा अनुसंधान संस्थान
INSTITUTE FOR PLASMA RESEARCH
परमाणु ऊर्जा विभाग, भारत सरकार का एक सहायता प्राप्त संस्थान
**An Aided Institute of Department of Atomic Energy,
Government of India**



इन्दिरा पुल के पास, भट, गांधीनगर - 382 428 भारत
दूरभाष: (079) 2396 2020/2021/2028
फैक्स: 91-079-23962277
वेब: www.ipr.res.in

NEAR INDIRA BRIDGE, BHAT
DIST. GANDHINAGAR - 382 428 (INDIA)
Phone: (079) 2396 2000/2026/2332
Fax : 91-079-23962277
Web : www.ipr.res.in

ENQUIRY

ENQUIRY NO : IPR/EQF/19-20/057
Date : 03-06-2019

Due on : 11-07-2019 by 1:00 PM IST

Please send your offer in sealed envelope specifying Enquiry No, Date & Due Date, ALONG WITH your credentials for the following items which we are interested to import directly against Foreign Trade Policy 2015-2020.

Important Note:

Please note that e-mail quotations are not acceptable however you may send your queries (if any) to importpurchase@ipr.res.in

Please ensure your sealed quotation reaches this office not later than above mentioned due date and time.

Kindly go through the following documents properly before quoting which are available on the IPR web portal i.e., http://www.ipr.res.in/documents/tender_terms.html / attached herewith.

- 1) Instructions to the bidders & Terms and conditions (refer Form No: **IPR-FP-01.V3**)
- 2) Bidding format

GST for Goods and Services (IGST/CGST/SGST TAX BENEFITS): Please refer **clause no: 14** of Form No: **IPR-FP-01.V3**

QUOTATION SHOULD BE ADDRESSED TO PURCHASE OFFICER ONLY

Sr No	Description	Quantity
1	26 pin D type electrical feed through for vaccum	4.0 Nos.

Note: 1. Please quote with complete technical details (Technical compliance sheet and product data sheet).
2. TDS as per CGST Act: As per provisions of section No. 51 of the CGST Act 2017, TDS @2% (IGST 2% or CGST 1% and SGST 1%) will be deducted while making payment to the suppliers where total value of orders/contracts/work orders exceeds Rs. 2.5 lakhs, in the event of order in Indian Rupees. Necessary TDS Certificate will be issued to the supplier after TDS deduction.

Encl: As per attached

Sd/-

Mr. D. Ramesh

Information to Vendors: We are working towards a single platform for our future requirement. Hence, please refer IPR website i.e, <http://www.ipr.res.in/documents/tendersenq.html> for our future requirement.

Specification of electrical feed through for vacuum applications

Quantity required = 4 Nos.

Specification	IPR Requirements
Feed through pin	26 pin D type male to female high density
Connection type	male to female high density (Air side Female and Vacuum side Male)
pin or contact material	copper alloy
Clip material	solder cup connection
Feed through shell	stainless steel shell
Pin current rating	≥ 5 Amp in vacuum per pin
Pin voltage rating	≥ 200 volt DC per pin
Temperature range of feed through	-20°C to 125°C, Vacuum side connector: -20°C to 150°C, Air side connector -20°C to 80°C
O -ring or sealing material	Viton
Leak Rate	≤5.0E-09 mbar-liter/sec

Compliance sheet of electrical feed through for vacuum applications

Quantity required = 4 Nos.

Specification	IPR Requirements	Vendors response
Feed through pin	26 pin D type male to female high density	
Connection type	male to female high density (Air side Female and Vacuum side Male)	
pin or contact material	copper alloy	
Clip material	solder cup connection	
Feed through shell	stainless steel shell	
Pin current rating	≥ 5 Amp in vacuum per pin	
Pin voltage rating	≥ 200 volt DC per pin	
Temperature range of feed through	-20°C to 125°C, Vacuum side connector: -20°C to 150°C, Air side connector -20°C to 80°C	
O -ring or sealing material	Viton	
Leak Rate	≤5.0E-09 mbar-liter/sec	

Bidder's Official Stamp and Sign:

Date: