



प्लाज्मा अनुसंधान संस्थान
INSTITUTE FOR PLASMA RESEARCH
परमाणु ऊर्जा विभाग, भारत सरकार का एक सहायता प्राप्त संस्थान
**An Aided Institute of Department of Atomic Energy,
Government of India**



इन्दिरा पुल के पास, भट, गांधीनगर - 382 428 भारत
दूरभाष: (079) 2396 2020/2021/2028
फैक्स: 91-079-23962277
वेब: www.ipr.res.in

NEAR INDIRA BRIDGE, BHAT
DIST. GANDHINAGAR - 382 428 (INDIA)
Phone: (079) 2396 2000/2026/2332
Fax : 91-079-23962277
Web : www.ipr.res.in

ENQUIRY

ENQUIRY NO : IPR/EQF/19-20/086
Date : 09-07-2019

Due on : 22-08-2019 by 1:00 PM IST

Please send your offer in sealed envelope specifying Enquiry No, Date & Due Date, ALONG WITH your credentials for the following items which we are interested to import directly against Foreign Trade Policy 2015-2020.

Important Note:

Please note that e-mail quotations are not acceptable however you may send your queries (if any) to importpurchase@ipr.res.in

Please ensure your sealed quotation reaches this office not later than above mentioned due date and time.

Kindly go through the following documents properly before quoting which are available on the IPR web portal i.e., http://www.ipr.res.in/documents/tender_terms.html / attached herewith.

- 1) Instructions to the bidders & Terms and conditions (refer Form No: **IPR-FP-01.V3**)
- 2) Bidding format

GST for Goods and Services (IGST/CGST/SGST TAX BENEFITS): Please refer **clause no: 14** of Form No: **IPR-FP-01.V3**

QUOTATION SHOULD BE ADDRESSED TO PURCHASE OFFICER ONLY

Sr No	Description	Quantity
1	Supply of Vacuum Guages with accessories.	6.0 Nos.

Note:

- 1.Delivery required within 12 to 14 weeks from the date of P.O
- 2.Please quote with complete technical details (Technical compliance sheet and product data sheet).
- 3.Vendor should supply the material at IPR Extension Lab. Gandhinagar.
- 4.TDS as per CGST Act: As per provisions of section No. 51 of the CGST Act 2017, TDS @2% (IGST 2% or CGST 1% and SGST 1%) will be deducted while making payment to the suppliers where total value of orders/contracts/work orders exceeds Rs. 2.5 lakhs, in the event of order in Indian Rupees. Necessary TDS Certificate will be issued to the supplier after

Encl: TDS deduction.
As per attached

Sd/-
Mr. D. Ramesh
Purchase Officer-II

Information to Vendors: We are working towards a single platform for our future requirement. Hence, please refer IPR website i.e,
<http://www.ipr.res.in/documents/tenderseng.html> for our future requirement.

Technical specifications for supply of vacuum gauge

- (a) Quantity : 06 Nos.
- (b) Measurement pressure range : 1×10^3 mbar to 1×10^{-4} mbar
- (c) Measurement Accuracy : $\leq \pm 15\%$
- (d) Operating temperature range : 5 C to 50 C
- (e) Overpressure limit : ≥ 4 bar
- (f) Inlet connections : DN 16 ISO KF
- (g) Operating voltage : 230 V \pm 10%, 50 Hz
- (h) Necessary accessories : Vacuum gauges should be supplied with suitable display and controller unit, connecting cables (minimum length required = 10 meter)

Acceptance criteria:

- (a) Vendor's representatives should demonstrate necessary test for vacuum gauges to show pressure range of 1×10^3 to 5×10^{-4} mbar as mentioned specification Sr. No. (b) at IPR site.
- (b) Vendor should provide operating manual along with final product delivery at IPR.

Packing and Delivery:

Vendor should take proper care and precaution in packing the items and delivering them safely to IPR Extension lab, Gandhinagar.

Compliance sheet for supply of Vacuum Gauges

Sr. No.	IPR Specifications		Vendor's specifications	Deviation (If any)
(a)	Quantity	06 Nos.		
(b)	Measurement pressure range	1×10^3 mbar to 1×10^{-4} mbar		
(c)	Measurement Accuracy	$\leq \pm 15\%$		
(d)	Operating temperature range	5 C to 50 C		
(e)	Overpressure limit	≥ 4 bar		
(f)	Inlet connections	DN 16 ISO KF		
(g)	Operating voltage	230 V \pm 10%, 50 Hz		
(h)	Necessary accessories	Vacuum gauges should be supplied with suitable display and controller unit, connecting cables (minimum length required = 10 meter)		
(i)	Acceptance criteria: (c) Vendor's representatives should demonstrate necessary test for vacuum gauges to show pressure range of 1×10^3 to 5×10^{-4} mbar as mentioned specification Sr. No. (b) at IPR site. (d) Vendor should provide operating manual along with final product delivery at IPR.			

Bidder's official stamp and sign:

Date: