

This file has been cleaned of potential threats.

To view the reconstructed contents, please SCROLL DOWN to next page.



प्लाज्मा अनुसंधान संस्थान
INSTITUTE FOR PLASMA RESEARCH
परमाणु ऊर्जा विभाग, भारत सरकार का एक सहायता प्राप्त
संस्थान



An Aided Institute of Department of Atomic Energy,
Government of India

इन्दिरा पुल के पास, भट, गांधीनगर - 382 428 भारत
दूरभाष: (079) 2396 2020/2021/2028
फैक्स: 91-079-23962277
वेब: www.ipr.res.in

NEAR INDIRA BRIDGE, BHAT
DIST. GANDHINAGAR - 382 428 (INDIA)
Phone: (079) 2396 2020/2021/2028
Fax : 91-079-23962277
Web : www.ipr.res.in

ENQUIRY

ENQUIRY NO : IPR/EQL/18-19/037
Date : 17-05-2018

Due on : 21-06-2018 by 1:00 PM IST

Reminder-1 Dt: 08-06-2018

Please send your offer in sealed envelope specifying Enquiry No, Date & Due Date, ALONG WITH your credentials for the following items:

Important Note:

Please note that e-mail quotations are not acceptable however you may send your queries (if any) to localpurchase@ipr.res.in

Please ensure your sealed quotation reaches this office not later than above mentioned due date and time.

Kindly go through the following documents properly before quoting which are available on the IPR web portal i.e., http://www.ipr.res.in/documents/tender_terms.html / attached herewith.

- 1) Instructions to the bidders & Terms and conditions (refer Form No: IPR-LP-01.V4)
- 2) Bidding format

GST for Goods and Services (IGST/CGST/SGST TAX BENEFITS): Please refer clause no: 8 of Form No: IPR-LP-01.V4

QUOTATION SHOULD BE ADDRESSED TO PURCHASE OFFICER ONLY

Sr No	Description	Quantity
1	Spark Gap	1.0 Nos.

Note: -

Encl: As per attachment.

Sd/-

Mr. D. Ramesh
Purchase Officer-II

Information to Vendors: We are working towards a single platform for our future requirement. Hence, please refer IPR website i.e, <http://www.ipr.res.in/documents/tenderseng.html> for our future requirement.

Specifications:

High Pressure Spark Gap

1. Breakdown Voltage: 10 KV DC
2. Max. Peak current: 50 KA
3. Terminal: M6
4. Size: Max. Diameter 90 mm
5. With suitable decoupling capacitor
6. Trigger voltage: 8 KV Pulse
7. Terminal: 3 Nos.

Compliance Form

Sr. No.	Parameters	Indenter requirement	Vendor Specification
01	Breakdown Voltage	10 KV DC	
02	Max. Peak current	50 KA	
03	Terminal	M6	
04	Size	Max. Diameter 90 mm	
05	Suitable decoupling capacitor	Yes	
06	Trigger voltage	8 KV Pulse	
07	Terminal	03 Nos.	

Bidder's sign with official stamp