

This file has been cleaned of potential threats.

To view the reconstructed contents, please SCROLL DOWN to next page.



प्लाज्मा अनुसंधान संस्थान
INSTITUTE FOR PLASMA RESEARCH
परमाणु ऊर्जा विभाग, भारत सरकार का एक सहायता प्राप्त
संस्थान
An Aided Institute of Department of Atomic Energy,
Government of India



इन्दिरा पुल के पास, भट, गांधीनगर - 382 428 भारत
दूरभाष: (079) 2396 2020/2021/2028
फैक्स: 91-079-23962277
वेब: www.ipr.res.in

NEAR INDIRA BRIDGE, BHAT
DIST. GANDHINAGAR - 382 428 (INDIA)
Phone: (079) 2396 2020/2021/2028
Fax : 91-079-23962277
Web : www.ipr.res.in

ENQUIRY

ENQUIRY NO : IPR/EQL/18-19/122
Date : 17-07-2018

Due on : 23-08-2018 by 1:00 PM IST

Please send your offer in sealed envelope specifying Enquiry No, Date & Due Date, ALONG WITH your credentials for the following items:

Important Note:

Please note that e-mail quotations are not acceptable however you may send your queries (if any) to localpurchase@ipr.res.in

Please ensure your sealed quotation reaches this office not later than above mentioned due date and time.

Kindly go through the following documents properly before quoting which are available on the IPR web portal i.e., http://www.ipr.res.in/documents/tender_terms.html / attached herewith.

- 1) Instructions to the bidders & Terms and conditions (refer Form No: **IPR-LP-01.V4**)
- 2) Bidding format

GST for Goods and Services (IGST/CGST/SGST TAX BENEFITS): Please refer **clause no: 8** of Form No: **IPR-LP-01.V4**

QUOTATION SHOULD BE ADDRESSED TO PURCHASE OFFICER ONLY

Sr No	Description	Quantity
1	Silver-Plated Multiple-Strands Copper-Cable, 120m long without joint, according to attached specifications	120.0 Mtrs.

Note: Please quote with complete technical details (Technical compliance sheet and product data sheet).

Encl: As per attachment.

Sd/-

Mr. D. Ramesh
Purchase Officer-II

Information to Vendors: We are working towards a single platform for our future requirement. Hence, please refer IPR website i.e, <http://www.ipr.res.in/documents/tenderseng.html> for our future requirement.

Silver-Plated Multiple-Strands Copper-Cable

Quantity = 1

Technical Specifications

No.	Particulars	Required Value
1	Cable conductor	Silver Plated Copper with multiple strands
2	Cable Insulation	PTFE plus Polyimide Insulated wire
3	Cable diameter	8.4 (mm) \pm 0.6 (mm) without insulation
4	Strand type	Multiple strands to maintain maximum flexibility
5	Insulation thickness	0.7 mm at max
6	Weight of cable	350 gm/meter at max
7	Cable resistance	<0.6 milli-Ohm/m
8	Temperature limitation	150°C
9	Cable length	120m without any joint
10	Cable sleeve color	RED

Silver-Plated Multiple-Strands Copper-Cable

Compliance Sheet

No.	Particulars	Required Value	Offer Value
1	Cable conductor	Silver Plated Copper with multiple strands	
2	Cable Insulation	PTFE plus Polyimide Insulated wire	
3	Cable diameter	8.4 (mm) \pm 0.6 (mm) without insulation	
4	Strand type	Multiple strands to maintain maximum flexibility	
5	Insulation thickness	0.7 mm at max	
6	Weight of cable	350 gm/meter at max	
7	Cable resistance	<0.6 milli-Ohm/m	
8	Temperature limitation	150°C	
9	Cable length	120m without any joint	
10	Cable sleeve color	RED	

Very Important Qualifying Conditions:

Bidders/Suppliers should write the values of the particulars explicitly in the column entitled 'Parameter to be specified in no./ value with Feature(s) of Quoted Item/s by Suppliers/Bidder.' Values or words like 'OK', 'Complied', 'Yes' or 'Will be provided' etc. will not be accepted and their quotations will out rightly be rejected