

This file has been cleaned of potential threats.

To view the reconstructed contents, please SCROLL DOWN to next page.



प्लाज्मा अनुसंधान संस्थान  
INSTITUTE FOR PLASMA RESEARCH  
परमाणु ऊर्जा विभाग, भारत सरकार का एक सहायता प्राप्त  
संस्थान  
An Aided Institute of Department of Atomic Energy,  
Government of India



इन्दिरा पुल के पास, भट, गांधीनगर - 382 428 भारत  
दूरभाष: (079) 2396 2020/2021/2028  
फैक्स: 91-079-23962277  
वेब: [www.ipr.res.in](http://www.ipr.res.in)

NEAR INDIRA BRIDGE, BHAT  
DIST. GANDHINAGAR - 382 428 (INDIA)  
Phone: (079) 2396 2020/2021/2028  
Fax : 91-079-23962277  
Web : [www.ipr.res.in](http://www.ipr.res.in)

## ENQUIRY

ENQUIRY NO : IPR/EQL/18-19/347  
Date : 09-01-2019

**Due on : 07-02-2019 by 1:00 PM IST**

Please send your offer in sealed envelope specifying Enquiry No, Date & Due Date, ALONG WITH your credentials for the following items:

### Important Note:

Please note that e-mail quotations are not acceptable however you may send your queries (if any) to [localpurchase@ipr.res.in](mailto:localpurchase@ipr.res.in)

Please ensure your sealed quotation reaches this office not later than above mentioned due date and time.

Kindly go through the following documents properly before quoting which are available on the IPR web portal i.e., [http://www.ipr.res.in/documents/tender\\_terms.html](http://www.ipr.res.in/documents/tender_terms.html) / attached herewith.

- 1) Instructions to the bidders & Terms and conditions (refer Form No: **IPR-LP-01.V4**)
- 2) Bidding format

**GST for Goods and Services (IGST/CGST/SGST TAX BENEFITS):** Please refer **clause no: 8** of Form No: **IPR-LP-01.V4**

### QUOTATION SHOULD BE ADDRESSED TO PURCHASE OFFICER ONLY

Sr No	Description	Quantity
1	3 Axis RF field strength meter	1.0 Nos.

Note: TDS as per CGST Act: As per provisions of section No. 51 of the CGST Act 2017, TDS @2% (IGST 2% or CGST 1% and SGST 1%) will be deducted while making payment to the suppliers where total value of orders/contracts/work orders exceeds Rs. 2.5 lakhs, in the event of order in Indian Rupees. Necessary TDS Certificate will be issued to the supplier after TDS deduction.

Encl: As per attachment.

Sd/-

Mr. D. Ramesh  
Purchase Officer-II

**Information to Vendors:** We are working towards a single platform for our future requirement. Hence, please refer IPR website i.e,

<http://www.ipr.res.in/documents/tenderseng.html> for our future requirement.

### 3 - Axis RF Field Strength Meter

#### Specifications

Measuring Range	: 38mV to 20 v/m
Frequency Range	: 10MHz - 8GHZ
Sensor	: Triple Axis ( X, Y, Z)
Manual Data Memory	: Max & AVG Records
Measurement Units	: mv/m, v/m, $\mu$ A/m , mA/m , $\mu$ W/m <sup>2</sup> , mW/m <sup>2</sup> , $\mu$ W/cm <sup>2</sup>
Display Resolution	: 0.1mv/m, 0.1 $\mu$ A/m, 0.001 $\mu$ W/m <sup>2</sup> , 0.001 $\mu$ W/cm <sup>2</sup>
Manual Data Memory	: Read storage -- 200 data sets
Alarm Function	. Adjustable
Calibration Factor CAL	: Adjustable threshold with ON/ OFF
Sampling Time	: 2.5 times per second
Operating Temp. & RH	: 0 to 50°C ( 32°F to 122°F ). Max. 80% RH.
Power Supply	: 9V battery /Quote
Battery life	: Quote
Weight	: Quote
Size	: Quote.
Accessories	: Instruction Manual and hard carrying case
Warranty	: 1 year
The vendor shall submit the test certificates	
Quantity	: 1 no.